2025/11/01 01:47 Thermoline 1315M Furnace

# **Thermoline 1315M Furnace**



Location: Microfluidics Lab

**Description:** Furnace

Manufacturer: Thermoline

#### **About**

"The furnace chamber is heated by a single three section resistant heater which is embedded in a refractory material. The chamber is insulated with a ceramic fiber insulation. The temperature is controlled by an electronic control. The temperature is measured by a thermocouple and is registered on a digital display. For safety, door switches are incorporated to remove power from the heating elements when the door is opened. The furnace is supported by the control section which also houses the electrical connections."

### **Safety Concerns**

# **Operating Procedures**

### **Detailed Specifications**

- Chamber dimensions: 4" W x 3.75" H x 4.5" D (10.2 x 9.5 x 11.4 cm)
- Temperature ranges: 100C-982C for continuous use, or 982C-1100C for intermittent use. Continuous use is operating the furnace for more than three straight hours, and intermittent

use is operating the furnace for less than three hours.

#### **Reference Documentation**

**Operations and Repair Manual** 

Manual

From

https://microfluidics.cnsi.ucsb.edu/wiki/ - Innovation Workshop Wiki

Permanent link:

https://microfluidics.cnsi.ucsb.edu/wiki/doku.php?id=thermoline\_furnace&rev=1672957595

Last update: 2023/01/05 22:26

